

Technical Information

Sterile BLM Liquid 5 (Ready to use)

Product Code: MS2314

Application: Sterile Beta Lactamase I Supplement is a highly purified enzyme based product specifically designed for the inactivation of Penicillins, thus finds its application in beta-lactams (Penicillins) sterility testing (either environmental or antibiotics sterility testing).

Composition**

Per vial sufficient for 20 ml medium

*Ingredients Concentration

Penicillinase Potency¹: 2000 Levy Units/ml/min

Inactivation of Pen G²:1,000,000 IU/ml/min

Note: 1: Levy Unit (LU): amount of Penicillinase that inactivates 593 IU of Sodium Penicillin G per minute at 25º C andpH 7.0, 2: Not all penicillins are comparable to Penicillin G in regards to their inactivation by Penicillinase. Some may require as much as 20 times more Penicillinase than does Penicillin G.

Directions

Sterile Beta lactamase I Supplement is an optimized ready to use solution that can be directly added to the test sample. The amount of product to be added to the test sample should be determined and set-up depending on the application, concentration of antibiotic to be inactivated and depending on the specific beta-lactam that has to be inactivated. Add appropriate amount of solution depending on the application. Remaining solution can be stored at 2-8°C. Aseptic techniques are to be followed throughout the procedure.

Type of specimen

Pharmaceutical samples

Specimen Collection and Handling

For pharmaceutical samples follow appropriate techniques for handling specimens as per established guidelines. After use, contaminated materials must be sterilized by autoclaving before discarding.

Warning & Precautions

For professional use only. Read the label before opening the container. Wear protective gloves/protective clothing/eye protection/face protection. Follow good microbiological lab practices while handling specimens and culture. Standard precautions as per established guidelines should be followed while handling clinical specimens. Safety guidelines may be referred in individual safety data sheets.

Disposal

User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with clinical sample must be decontaminated and disposed of in accordance with current laboratory techniques (1, 2).

Storage and Shelf Life

Store at 2-8°C. Use before expiry date on the label.



Reference

- 1. Isenberg (Ed.), 2004, Clinical Microbiology Procedures Handbook, Vol.3, American Society for Microbiology, Washington. D.C.
- 2. Jorgensen, J.H., Pfaller, M.A., Carroll, K.C., Funke, G., Landry, M.L., Richter, S.S and Warnock., D.W. (2015) Manual of Clinical Microbiology, 11th Edition. Vol. 1.

* Not for Medicinal Use

Disclaimer

- User must ensure suitability of the product(s) in their application prior to use.
- The product conforms solely to the technical information provided in this booklet and to the best of knowledge research and development work carried at **CDH** is true and accurate.
- Central Drug House Pvt. Ltd. reserves the right to make changes to specifications and information related to the products at any time.
- Products are not intended for human or animal diagnostic or therapeutic use but for laboratory, research or further manufacturing of diagnostic reagents extra.
- Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.
- Do not use the products if it fails to meet specifications for identity and performance parameters.