



Ready Prepared Media

Technical Information

R-2A Agar Plate

Product Code: PM 1962

Application: Recommended for heterotrophic plate count of water samples using longer incubation periods.

Composition**

Ingredients	Gms / Litre
Acicase#	0.500
Yeast extract	0.500
Proteose peptone	0.500
Dextrose (Glucose)	0.500
Starch soluble	0.500
Dipotassium hydrogen phosphate	0.300
Magnesium sulphate	0.024
Sodium pyruvate	0.300
Agar	15.000
Final pH (at 25°C)	7.2±0.2

**Formula adjusted, standardized to suit performance parameters

Equivalent to Casein Acid Hydrolysate

Principle & Interpretation

The heterotrophic plate count (HPC), formerly known as the standard plate count is a procedure for estimating the number of live heterotrophic bacteria in water and measuring changes during water treatment, in distribution systems or in swimming pools. R-2A Agar is recommended by APHA (1,2) for estimating the heterotrophic plate count by the pour plate, spread plate or membrane filter procedure. R-2A Agar is formulated as per Reasoner and Geldreich (3). Stressed or injured organisms during water treatment are unable to grow on high nutrient media, since the faster growing organisms outgrow the former (4). Therefore, the use of a low nutrient medium like R-2A Agar incubated for longer incubation periods allows these stressed organisms to grow well. Many bacteria from natural waters which contain limited nutrients at ambient temperature, grow best on the media with less nutrient levels. They grow better at the temperatures below the routine laboratory incubation temperatures of 35 to 37°C (4). Acicase, proteose peptone and yeast extract provide nitrogen, carbon compounds, vitamins, amino acids and minerals. Dextrose/ glucose serves as an energy source. Soluble starch aids in the recovery of injured organisms by absorbing toxic metabolic byproducts while sodium pyruvate increases the recovery of stressed cells. Magnesium sulphate is a source of divalent cations and sulphate. Dipotassium hydrogen phosphate is used to balance the pH of the medium. The number of colonies on a plate are reported as CFU (Colony Forming Units) per volume of sample.

Type of specimen

Water samples

Specimen Collection and Handling:

For water samples, follow appropriate techniques for sample collection, processing as per guidelines and local standard (1). After use, contaminated materials must be sterilized by autoclaving before discarding.

Warning and Precautions :

Read the label before opening the pack. Wear protective gloves/protective clothing/eye protection/face protection. Follow good microbiological lab practices while handling specimens and culture. Standard precautions as per established guidelines should be followed while handling specimens. Safety guidelines may be referred in individual safety data sheets.

Limitations :



Central Drug House (P) Ltd. Corp. Office : 7/28, Vardaan House, Ansari Road, Daryaganj, New Delhi-110002 (INDIA)

Phone : +91-11-49404040 (100 Lines) Fax : +91-11-49404050 E-mail : sales@cdhfinechemical.com Website : www.cdhfinechemical.com



Ready Prepared Media

1. Individual organisms differ in their growth requirement and may show variable growth patterns on the medium
2. Each lot of the medium has been tested for the organisms specified on the COA. It is recommended to users to validate the medium for any specific microorganism other than mentioned in the COA based on the user's unique requirement.
3. Longer incubation time other than specified is required for slow growing microorganisms.
4. The media is intended for water samples for recovery of stressed or injured organisms.
5. It is recommended to store the plates at 24-30°C to avoid minimum condensation.

Performance and Evaluation

Performance of the medium is expected when used as per the direction on the label within the expiry period when stored at recommended temperature.

Methodology

Either streak, inoculate or surface spread the test inoculum (50-100 CFU) aseptically on the plate.

Quality Control

Appearance

Sterile R-2A Agar in 90 mm disposable plates.

Colour of medium

Light yellow coloured medium

Quantity of medium

25 ml of medium in 90 mm disposable plates.

pH

7.00-7.40

Sterility Check

Passes release criteria

Cultural Response

Cultural characteristics observed *by using standard ATCC cultures after an incubation at 30-35°C for 24-72 hours. (*-In case of water samples from fields it is suggested to incubate further for upto 7 days to ascertain the absence of organisms)

Cultural Response

Organism	Inoculum (CFU)	Growth	Recovery
Candida albicans ATCC 10231 (00054*)	50-100	good-luxuriant	>=50%
Escherichia coli ATCC 25922 (00013*)	50-100	good-luxuriant	>=50%
Salmonella Enteritidis ATCC 13076 (00030*)	50-100	good-luxuriant	>=50%
Pseudomonas aeruginosa ATCC 9027 (00026*)	50-100	good-luxuriant	>=50%
Staphylococcus aureus subsp. aureus ATCC 6538 (00032*)	50-100	good-luxuriant	>=50%
Bacillus subtilis subsp. spizizenii ATCC 6633 (00003*)	50-100	good-luxuriant	>=50%
Aspergillus brasiliensis ATCC 16404 (00053*)	50-100	good-luxuriant	>=50%
Enterococcus faecalis ATCC 29212 (00087*)	50-100	good-luxuriant	>=50%
Salmonella Typhi ATCC 6539	50-100	good-luxuriant	>=50%

Key : * Corresponding WDCM numbers.





Ready Prepared Media

Storage and Shelf Life

On receipt store between 20-30°C. Use before expiry date on the label. Product performance is best if used within stated expiry period.

Disposal

User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with sample must be decontaminated and disposed of in accordance with current laboratory techniques (5,6).

Further Reading

1. Lipps WC, Braun-Howland EB, Baxter TE, eds. Standard methods for the Examination of Water and Wastewater, 24th ed. Washington DC:APHA Press; 2023.
2. Salfinger Y., and Tortorello M.L. Fi015, Compendium of Methods for the Microbiological Examination of Foods, 5th Ed., American Public Health Association, Washington, D.C.
3. Reasoner D. J. and Geldreich E. E., 1985, Appl. Environ. Microbiol., 49:1.
4. Collins V. J. and Willoughby J. G., 1962, Arch. Microbiol., 43:294.
5. Isenberg, H.D. Clinical Microbiology Procedures Handbook. 2nd Edition.
6. Jorgensen, J.H., Pfaller, M.A., Carroll, K.C., Funke, G., Landry, M.L., Richter, S.S and Warnock, D.W. (2015), Manual of Clinical Microbiology, 11th Edition. Vol. 1.66

Disclaimer :

- User must ensure suitability of the product(s) in their application prior to use.
- The product conform solely to the technical information provided in this booklet and to the best of knowledge research and development work carried at **CDH** is true and accurate
- **Central Drug House Pvt. Ltd.** reserves the right to make changes to specifications and information related to the products at any time.
- Products are not intended for human or animal diagnostic or therapeutic use but for laboratory, research or further manufacturing of diagnostic reagents extra.
- Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.
- Do not use the products if it fails to meet specifications for identity and performance parameters.

