



Product Specification

cdhfinechemical.com

CONDUCTIVITY STANDARD 12880 CRISTAR® MICROSIEMENSE IN ACCORDANCE WITH NIST

PRODUCT CODE	818924
---------------------	---------------

Intended Use

Standard Solution for conductivity 12880 µS/cm is used as standard solution for measurement of conductivity in calibration.

Principle And Interpretation

A conductivity standard is an electrolyte solution with an accurately known electrolytic conductivity. Accuracy of conductivity depends on the conductivity measuring equipment and the application of appropriate standards.

PARAMETER	LIMIT
Description	Clear colourless solution
Solubility	Miscible with water
Conductivity at 25°C	12880.0 ± 1.0% µS/Cm

Note(s) : Assay (if applicable) method mentioned.

<p>WARNING</p> <p>Hazard statements: Not hazardous. No hazards.</p> <p>Precautionary statements</p> <p>Prevention: -----</p> <p>Response: -----</p> <p>Disposal:-----</p>	<p>IMDG Code : --</p> <p>UN No. : --</p> <p>IATA : --</p>
---	---

Hazard Pictogram(s) :-----

Replace Date 30-Dec-2023