

Technical Information

MiViral™ Transport Kit (D)

Product Code: MS 3760D

Application: - w/ 3 mL Viral Transport Medium in a 10 mL self-standing tube and one sterile viscose swab without breakpoint

Introduction:

MiMedia's MiViral™ Transport Kit (D) is a specially designed transport system to collect and transport viruses in an active form to the laboratory for isolation. It is designed to maintain the viability and the virulence of the viral sample.

MiViral™ Transport Medium is made of Hanks Balanced Salt Solution and contains a protective protein, antibiotics to control microbial and fungal contamination and buffers to control the pH. Phenol red is used as a pH indicator. The medium also contains a cryoprotectant which helps in preserving the viruses if specimens are frozen for prolonged storage.

The viscose swab is a tipped swab on a polypropylene shaft, which can be used for collecting throat samples. It does not have a breakpoint.

Kit Contents:

Code	Description	Quantity	Storage
AL1167	MiViral™ Transport Medium	50 x 3ml	RT
PW1043	Sterile viscose swab w/o breakpoint	50 nos	RT

Procedure:

A. Collection of Samples

For a complete diagnostic analysis of viral diseases, it is important that the infectivity of the viruses is preserved after sample collection. The infectivity of viruses' decreases over time and the decay rate is generally a function of temperature. Stability of samples is enhanced by cooling therefore samples should be kept at 2-8°C. The probability of a successful isolation is more if the samples are processed immediately after collection and the viral load in the sample is more. Viral load is maximum if the samples are collected immediately after the onset of clinical symptoms and before the administration of antiviral medications.

B. Directions:

1. Cut open the pouch to remove the swab.
2. Specimen can be collected with the swab in the following manner.

Nasal swab

Nasal swab is collected for anterior turbinate. Insert dry swab into nostril and leave in place for a few seconds. Slowly withdraw it with a rotating motion.

Nasopharyngeal swab

Insert dry swab in to nostril and back to the nasopharynx. Leave in place for a few seconds. Slowly withdraw the swab with a rotating motion.

Throat swab

Ask patient to open his/her mouth. Swab the back of the throat near the tonsils thoroughly.

3. Break the swab near the break point and insert swab into the tube containing viral transport medium and close the cap tightly.
4. Label the sample correctly with the name of the patient and time and date of collection.
5. Transport the samples immediately to the laboratory for processing

Transportation of the Samples:

Samples should be transported to the laboratory as soon as possible.

Samples can be refrigerated at 2-8°C after collection or can be transported at 2-8°C on wet ice within 48 hours. If a long delay is expected in transit and processing, samples should be transported on dry ice and should be frozen at -70°C.



Dehydrated Culture Media
Bases / Media Supplements

Precautions:

1. Isolation of viruses will largely depend on proper specimen collection, timing of sample collection and processing of samples.
2. Specimen collection should be done in the acute phase of illness.
3. Do not use the product if (1) there is change in the color of the medium, (2) there is evidence of leakage, (3) there are other signs of deterioration.
4. To maintain infectivity of viruses it is important that temperature be properly maintained for sample collection to processing.
5. Avoid repeated freeze-thaw of collected samples.
6. It is recommended to refer to standard procedures and published protocols for sample collection and processing

Quality Control:**Appearance**

Orange coloured clear solution

pH at 25°C

7.3 ± 0.3

Osmolality in mOsm/Kg H₂O

500.00 - 600.00

Sterility

No bacterial or fungal growth is observed after 14 days of incubation as per USP specification.

Storage and Shelf Life

Store at 15-30°C.

Use before expiry date given on the product label.

Disclaimer :

- User must ensure suitability of the product(s) in their application prior to use.
- The product conforms solely to the technical information provided in this booklet and to the best of knowledge research and development work carried at **CDH** is true and accurate.
- **Central Drug House Pvt. Ltd.** reserves the right to make changes to specifications and information related to the products at any time.
- Products are not intended for human or animal diagnostic or therapeutic use but for laboratory, research or further manufacturing of diagnostic reagents extra.
- Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.
- Do not use the products if it fails to meet specifications for identity and performance parameters.



Central Drug House (P) Ltd. Corp. Office : 7/28, Vardaan House, Ansari Road, Daryaganj, New Delhi-110002 (INDIA)

Phone : +91-11-49404040 (100 Lines) Fax : +91-11-49404050 E-mail : sales@cdhfinechemical.com Website : www.cdhfinechemical.com