

# **Technical Information**

### Costein's LDS Test Medium

Product Code: DM 2621

**Application:** - Costeins LDS Test Medium is used for the identification of members of *Enterobacteriaceae* on the basis of lysine decarboxylase and hydrogen sulphide production.

## Composition\*\*

high degree of reliability (3).

| Composition                                       |                |             |  |  |
|---|----------------|-------------|--|--|
| Ingredients                                       | Gms / Litre    | Gms / Litre |  |  |
| Meat peptone                                      | 4.500          |             |  |  |
| Papaic digest of soyabean meal                    | 2.000          |             |  |  |
| Yeast extract                                     | 3.000          |             |  |  |
| Sodium chloride                                   | 5.000          |             |  |  |
| D-Glucose   | 1.000          |             |  |  |
| L-Lysine monohydrochloride                        | 10.000         |             |  |  |
| Sosium thiosulphate                               | 0.200          |             |  |  |
| Ammonium iron (II) sulphate                       | 0.200          |             |  |  |
| Bromocresol purple                                | 0.032          |             |  |  |
| Agar  | 6.000          |             |  |  |
| Final pH ( at 25°C)                               | 5.6±0.2        |             |  |  |
| **Formula adjusted, standardized to suit performa | nce parameters |             |  |  |

Torritata dajastea, startaaraizea to sait performance pa

Principle & Interpretation

Costeins LDS (Lysine Decarboxylase Sulfhydrase) Test Medium is formulated by Costein (1). This is the test culture medium for the simultaneous detection of lysine decarboxylase (LDC) and hydrogen sulphide production (2). Pietzsch used this medium on account of its

Meat peptone, papaic digest of soyabean meal and yeast extract in the medium supply nitrogen and other nutrients necessary to support bacterial growth. Sodium chloride helps to maintaining osmotic balance. Bromocresol purple is the pH indicator. Sodium thiosulphate serves as a reducing agent and helps to maintain a low oxygen tension in the medium. *Enterobacteriaceae* grows poorly at low pH; therefore, their growth is poor on this medium due to its low pH value of 5.6. LDC-positive species tend to neutralize the medium as a result of cadaverine production due to decarboxylation of lysine; the conditions for growth are thus improved and the pH indicator changes its colour from yellow to violet. Species which can also reduce thiosulphate to hydrogen sulphide, cause an additional blackening of the violet medium due to the precipitation of iron sulphide. LDC-negative species do not increase the pH value of the medium; the pH indicator does not undergo a colour change. Therefore growth of these microorganisms is poor and H<sub>2</sub>S positive species are thus unable to produce any hydrogen sulphide e.g. *Citrobacter, Proteus vulgaris, Proteus mirabilis, Providentia, Enterobacter, Shigella* show yellow coloured growth. LDC-positive and H<sub>2</sub>S positive organisms give black colour and may be surrounded by a violet zone e.g. Salmonella, Edwardsiella and others LDC-positive and H<sub>2</sub>S negative organism are violet in colour e.g. *Escherichia, Klebsiella, Hafnia, Serratia* and some rare *Salmonella* and others.

## Methodology

Suspend 32 grams of dehydrated media in 1000 ml distilled water. Mix thoroughly & dispense into test tubes to a depth of approximately 5 cm, if possible cover with a 5 mm layer of viscous paraffin. Sterilize by autoclaving at 15 lbs pressure (121°C) for 15 minutes. Allow the tubes to solidify in a vertical position.





## **Quality Control**

#### **Appearance**

Cream to yellow homogeneous free flowing powder

#### Gelling

Firm, comparable with 0.6 % Agar gel

#### **Colour and Clarity**

Yellow coloured, clear to slightly opalescent gel forms in tubes as slants

#### Reaction

Reaction of 3.2% w/v agueous solution at 25°C. pH: 5.6±0.2

#### pH Range

5.40-5.80

### **Cultural Response**

DM2621: Cultural characteristics observed after an incubation at 35-37°C for 24-48 hours.

| Organism                          | Inoculum<br>(CFU) | Growth | Colour change of medium |
|-----------------------------------|-------------------|--------|-------------------------|
| Cultural Response                 |                   |        |                         |
| Escherichia coli ATCC 25922       | 50-100            | good   | violet                  |
| Serratia marcescens ATCC 14756    | 50-100            | good   | violet                  |
| Salmonella Typhimurium ATCC 14028 | 50-100            | good   | violet and black        |
| Salmonella Enteritidis ATCC 13076 | 50-100            | good   | violet and black        |
| Salmonella Arizonae ATCC 13314    | 50-100            | good   | violet and black        |
| Shigella flexneri ATCC 12022      | 50-100            | fair   | yellow                  |
| Citrobacter freundii ATCC 8090    | 50-100            | fair   | yellow                  |
| Proteus mirabilis ATCC 29906      | 50-100            | fair   | yellow                  |
| Proteus vulgaris ATCC 13315       | 50-100            | fair   | yellow                  |

# Storage and Shelf Life

**Dried Media:** Store below 30°C in tightly closed container and the prepared medium at 2 - 8°C. Use before expiry date on the label. **Prepared Media:** 2-8° in sealable plastic bags for 2-5 days.

# **Further Reading**

- 1. Costein I. D., 1968, Einzelnahrboden fur die biochemische Ausscheidung van Salmonella-und Arizona-Kulturen- Zbl. F. Bakt. I. Orig., 206:390-395.
- 2. Edwards P. R., Fife M. A., 1961, Lysine-Iron Agar in the detection of Arizona cultures, Appl. Microbiol., 9:478-480.
- 3. Pietzch O., 1975, Der Voges-Proskauer- Schnelltest und der Lysindecarboxylase-Sulfhydrase-Test, Zwei Schnellmethoden fur die Enterobacteriaceae Diagnostik., Arch. lebensmittelhyg., 26: 23-24.

### Disclaimer:

- User must ensure suitability of the product(s) in their application prior to use.
- The product conform solely to the technical information provided in this booklet and to the best of knowledge research and development work carried at CDH is true and accurate
- Central Drug House Pvt. Ltd. reserves the right to make changes to specifications and information related to the products at any time.
- Products are not intended for human or animal diagnostic or therapeutic use but for laboratory, research or further manufacturing of diagnostic reagents extra.
- Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents. Do not use the products if it fails to meet specifications for identity and performances parameters.

