



# Product Specification

cdhfinechemical.com

## CLENTORN™ MA 02 NEUTRAL FOR CLEANING LABORATORY GLASSWARE INSTRUMENTS

<b>PRODUCT CODE</b>	<b>D1100</b>
<b>SYNONYMS</b>	--
<b>C.I. NO.</b>	--
<b>CASR NO.</b>	--
<b>ATOMIC OR MOLECULAR FORMULA</b>	--
<b>ATOMIC OR MOLECULAR WEIGHT</b>	--
<b>PROPERTIES</b>	--
<b>PARAMETER</b>	<b>LIMIT</b>
Description	Colourless to pale yellow, almost clear liquid.
Solubility	Miscible with water forming foams.
pH of (ready to use)	6.0-7.5
Wt. per ml at 20°C	About. 1.02 g
Viscosity	1500 cPs

For cleaning of equipment made of alkalisensitive metals (aluminium, zinc)and alloys with the same properties, as well as quartz and glass precision measuring equipment. Contains anionic and nonionic surfactants, phosphates and small quantities of adjuvants. depending on degree of soiling. Completely immerse the objects being cleaned in the aqueous solution. Increasing the bath temperature accelerates the cleaning process. Cleaning period : 2-24 hours depending on degree of soiling. Rinse with tap water immediately after removing from bath, and then rinse with demineralised water.  
Dry in air or in drying cupboard.

**Note(s) : Assay (if applicable) method mentioned.**

<b>WARNING</b> <b>Hazard statements:</b> Not hazardous. No hazards. <b>Precautionary statements</b> <b>Prevention:</b> ----- <b>Response:</b> ----- <b>Disposal:</b> -----	IMDG Code : -- UN No. : -- IATA : --
---	--

**Hazard Pictogram(s) :**-----

Replace Date 29-May-2023