

## Technical Information

### Columbia Blood Agar Base w/ 1% Agar

#### Product Code: DM 1144A

**Application:** - Columbia Blood Agar Base is used as an efficient base for preparation of blood agar, chocolate agar and for various selective and identification media.

#### Composition\*\*

Ingredients	Gms / Litre
Peptone, special	23.000
Corn starch	1.000
Sodium chloride	5.000
Agar	10.000
Final pH ( at 25°C)	7.3±0.2

\*\*Formula adjusted, standardized to suit performance parameters

#### Principle & Interpretation

Columbia Blood Agar Base is a general-purpose nutritious agar base devised by Ellner et al <sup>(1)</sup> in 1966 which is further enriched by the addition of sterile blood. Earlier blood agar bases contain either casein hydrolysate to give rapid production of large colonies or meat infusion to give defined hemolytic reactions. Columbia Agar Base contain both to give an improved all round performance. This medium with the added special peptone supports rapid and luxuriant growth of fastidious and nonfastidious organisms promotes typical colonial morphology; better pigment production and more sharply defined haemolytic reactions. Columbia Agar Base is used as the base for the media containing blood and selective formulations in different combinations of antimicrobial agents for isolation of different type of pathogens. Fildes found that Nutrient Agar supplemented with a digest of sheep blood supplied both of these factors and the medium would support the growth of *H. influenzae* <sup>(2, 3)</sup>. The inclusion of bacitracin makes the enriched Columbia Agar Medium selective for the isolation of *Haemophilus* species from clinical specimens, especially from upper respiratory tract <sup>(4)</sup>.

Corn starch serves as an energy source and also neutralizes toxic metabolites. Sheep blood permits the detection of haemolysis and also provides heme (X factor) which is required for the growth of many bacteria. However it is devoid of V factor (Nicotinamide adenine dinucleotide) and hence *Haemophilus influenzae* which needs both the X and V factors, will not grow on this medium. As this medium has a relatively high carbohydrate content, beta-haemolytic Streptococci may exhibit a greenish haemolytic reaction which may be mistaken for the alpha haemolysis.

Columbia Agar Base with added sterile serum provides an efficient medium for *Corynebacterium diphtheriae* virulence test medium. In which lines of toxin-antitoxin precipitation are clearly visible in 48 hours. Many pathogens require carbon dioxide; therefore, plates may be incubated in an atmosphere containing approximately 3-10% CO<sub>2</sub>.

Precaution: *Brucella* cultures are highly infective and must be handled carefully; incubate in 5-10% CO<sub>2</sub>. *Campylobacter* species are best grown at 42°C in a microaerophilic atmosphere. Plates with Gardnerella supplements should be incubated at 35°C for 48 hours containing 7% CO<sub>2</sub> <sup>(5)</sup>.

#### Methodology

Suspend 39 grams of powder media in 1000 ml distilled water. Shake well and heat to dissolve the medium completely. Sterilize by autoclaving at 15 lbs pressure (121°C) for 15 minutes. Cool to 45-50°C before adding heat sensitive compounds.

For Blood Agar: Add 5% v/v sterile defibrinated sheep blood to sterile cool base.

For Chocolate Agar: Add 10% v/v sterile defibrinated sheep blood to sterile cool base. Heat to 80°C for 10 minutes with constant agitation. The medium can be made selective by adding different antimicrobials to sterile base.

For Brucella species: Add rehydrated contents of 1 vial of Brucella Selective Supplement (MS2005) to 500 ml sterile molten base.

For Campylobacter species: Add rehydrated contents of 1 vial of Campylobacter Supplement - I (Blaser-Wang) (MS2006) or Campylobacter Supplement - II, (Butzler) (MS2007) or Campylobacter Supplement - III (Skirrow) (MS2008) or Campylobacter Selective Supplement (MS2090) or Campylobacter Supplement - VI (Butzler) (MS2106) to 500 ml sterile molten base along with rehydrated contents of 1 vial of Campylobacter Growth Supplement (MS2009).

For Gardnerella species: Add rehydrated contents of 1 vial of G.Vaginalis Selective Supplement (MS2056) to 500 ml sterile molten base.

For Cocci: Add rehydrated contents of 1 vial of Staph-Strepto Supplement (FD030) or Strepto Supplement (MS2031) or Streptococcus Selective Supplement (MS119) to 500 ml sterile molten base.

## Quality Control

### Physical Appearance

Cream to yellow homogeneous free flowing powder

### Gelling

Firm, comparable with 1.0% Agar gel.

### Colour and Clarity of prepared medium

Basal medium: Light amber coloured clear to slightly opalescent gel. After addition of 5%v/v sterile defibrinated blood : Cherry red coloured opaque gel forms in Petri plates

### Reaction

Reaction of 3.9% w/v aqueous solution at 25°C. pH : 7.3±0.2

**pH range** 7.10-7.50

### Culture Response/Characteristics

DM: 1144A Culture response absolved with added 5% w/v sterile delibrinated blood after on incubation of 35-37° C to 24-48 hours.

Organism	Inoculum (CFU)	Growth	Recovery	Haemolysis
Neisseria meningitidis ATCC 13090	50-100	luxuriant	>=70%	none
Staphylococcus aureus ATCC 25923	50-100	luxuriant	>=70%	beta / gamma
Staphylococcus aureus ATCC 6538	50-100	luxuriant	>=70%	beta / gamma
Staphylococcus epidermidis ATCC 12228	50-100	luxuriant	>=70%	gamma
Streptococcus pneumoniae ATCC 6303	50-100	luxuriant	>=70%	alpha
Streptococcus pyogenes ATCC 19615	50-100	luxuriant	>=70%	beta

## Storage and Shelf Life

**Dried Media:** Store below 30°C in tightly closed container and use before expiry date as mentioned on the label.

**Prepared Media:** 2-8° in sealable plastic bags for 2-5 days.

## Further Reading

1. Ellner P. P., Stoessel C. J., Drakeford E. and Vasi F., 1966, Am. J. Clin. Pathol., 45:502.
2. Fildes P., 1920, Br. J. Exp. Pathol., 1:129.
3. Fildes P., 1921, Br. J. Exp. Pathol., 2:16.
4. Chapin K. C. and Doern G. V., 1983, J. Clin. Microbiol., 17:1163.
5. Bailey R. K., Voss J. L. and Smith R. F., 1979, J. Clin. Microbiol., 9 ; 65-71

## Disclaimer :

- User must ensure suitability of the product(s) in their application prior to use.
- The product conform solely to the technical information provided in this booklet and to the best of knowledge research and development work carried at **CDH** is true and accurate
- **Central Drug House Pvt. Ltd.** reserves the right to make changes to specifications and information related to the products at any time.
- Products are not intended for human or animal diagnostic or therapeutic use but for laboratory, research or further manufacturing of diagnostic reagents extra.
- Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.
- Do not use the products if it fails to meet specifications for identity and performance parameters.