

Technical Information

KF Streptococcus Agar Base w/ BCP

Product Code: DM 2007

Application: - KF Streptococcus Agar Base w/ BCP is recommended for detection and enumeration of faecal Streptococci.

Composition**

Ingredients	Gms / Litre
Proteose peptone	10.000
Yeast extract	10.000
Sodium chloride	5.000
Sodium glycerophosphate	10.000
Maltose	20.000
Lactose	1.000
Sodium azide	0.400
Bromocresol purple	0.015
Agar	20.000
Final pH (25°C)	7.2±0.2

**Formula adjusted, standardized to suit performance parameters

Principle & Interpretation

Streptococci are gram-positive cocci and form a part of the normal commensal flora of the mouth, skin, intestine, upper respiratory tract of humans. Streptococci found in the faeces is Streptococci with group D Lancefield antigens. The types include *Streptococcus faecalis*, *Streptococcus faecium*, *Streptococcus bovis* and *Streptococcus duran*. They are low-grade pathogens and rarely cause disease. However, they may cause urinary tract infection in catheterized patients; mixed abdominal wound infections following gut surgery; and endocarditis on abnormal valves. Kenner - Faecal (KF) Medium were developed by Kenner et al^(1, 2) for isolating Streptococci in water and food materials.

Proteose peptone along with yeast extract provides nitrogen, carbon, sulphur, amino acids, vitamins and trace ingredients to the faecal Streptococci. Lactose and maltose are the fermentable carbohydrates and therefore serve as energy sources. Sodium azide is a selective agent, which hampers the growth of many gram-negative bacteria.

2,3,5-Triphenyl Tetrazolium Chloride is reduced to insoluble formazan by actively metabolizing cells, resulting in the formation of pink or red colonies. Bacteria resistant to azide, utilize lactose and / or maltose. The acidity so produced changes the colour of the indicator dyes to yellow.

Samples can be directly streaked or sterile membrane filters through which the water samples have been passed are aseptically placed on the media. After an incubation at 35-37°C for 24-48 hours, Enterococci appear as pink to red colonies. After this presumptive identification, further confirmatory tests should be carried out^(3, 4) using standard procedure.

Methodology

Suspend 76.41 grams of powder media in 1000 ml distilled water. Shake well & heat to dissolve the medium completely. DO NOT AUTOCLAVE. Overheating will lower the pH and render the medium less productive. Cool to 50°C and aseptically add 10 ml of 1% 2, 3, 5-Triphenyl Tetrazolium Chloride (TTC) (MS2057) to sterile medium. Mix well and pour into sterile Petri plates.

Quality Control

Physical Appearance

Cream to greyish yellow homogeneous free flowing powder

Gelling

Firm, comparable with 2.0% agar gel.

Colour and Clarity of prepared medium

Light purple coloured, clear to slightly opalescent gel forms in Petri plates

Reaction

Reaction of 7.64% w/v aqueous solution at 25°C. pH : 7.2±0.2

pH Range 7.00-7.40

Cultural Response/ characteristics

DM2007: Cultural characteristics observed with added MS2057, after an incubation at 35-37°C for 48-72 hours.

Organism	Inoculum (CFU)	Growth	Recovery	Colour of colony
<i>Enterobacter aerogenes</i> ATCC 19048	$\geq 10^3$	inhibited	0%	
<i>Enterococcus faecalis</i> ATCC 29212	50-100	good-luxuriant	$\geq 50\%$	red-maroon
<i>Escherichia coli</i> ATCC 25922	$\geq 10^3$	inhibited	0%	

Storage and Shelf Life

Dried Media: Store below 30°C in tightly closed container and use before expiry date as mentioned on the label.

Prepared Media: 2-8° in sealable plastic bags for 2-5 days.

Further Reading

1. Kenner B. A., Clark H. F. and Kabler P. W., 1960, Am. J. Public Health, 50:1553.
2. Kenner B. A., Clark H. F. and Kabler P. W., 1961, Appl. Microbiol., 9:15.
3. MacFaddin J. F., 1985, Media for Isolation-Cultivation-Identification-Maintenance of Medical Bacteria, Vol. 1, Williams and Wilkins, Baltimore.
4. Facklam R. R. and Moody M. P., 1970, Appl. Microbiol., 20:245.

Disclaimer :

- User must ensure suitability of the product(s) in their application prior to use.
- The product conform solely to the technical information provided in this booklet and to the best of knowledge research and development work carried at **CDH** is true and accurate
- **Central Drug House Pvt. Ltd.** reserves the right to make changes to specifications and information related to the products at any time.
- Products are not intended for human or animal diagnostic or therapeutic use but for laboratory, research or further manufacturing of diagnostic reagents extra.
- Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.
- Do not use the products if it fails to meet specifications for identity and performance parameters.