

Technical Information

Endo Agar with NaCl

Product Code: DM 2258

Application: - Endo Agar with NaCl is used for detection and isolation of pathogenic enteric bacilli.

Composition**

Ingredients	Gms / Litre
Peptone, special	8.000
Lactose	10.000
Sodium chloride	3.000
Dipotassium phosphate	2.000
Sodium sulphite	2.500
Basic fuchsin	0.200
Agar	12.000
Final pH (at 25°C)	7.5±0.2

**Formula adjusted, standardized to suit performance parameters

Principle & Interpretation

Endo Agar was developed by Endo to differentiate gram-negative bacteria on the basis of lactose fermentation, while inhibiting gram-positive bacteria ⁽¹⁾ by the incorporation of sodium sulphite and basic fuchsin. Endo Agar w/ NaCl is prescribed in the regulations for the execution of the German Meat Inspection Law ⁽²⁾.

The medium contains peptone special which provides nitrogen, carbon, vitamins and minerals required for bacterial growth. Sodium sulphite and basic fuchsin inhibits most of the gram-positive bacteria. Lactose fermenting *Escherichia coli* and coliforms produce aldehyde and acid. The aldehyde liberates fuchsin from the fuchsin-sulphite complex and colonies of lactose fermenters appear dark red. Non-lactose fermenters show colourless colonies.

With *Escherichia coli*, this reaction is very pronounced as the fuchsin crystallizes, exhibiting a permanent greenish metallic luster (fuchsin luster) to the colonies. Medium should be stored away from light to avoid photo-oxidation.

Methodology

Suspend 37.7 grams of powder media in 1000 ml distilled water. Shake well & heat to dissolve the medium completely. Sterilize by autoclaving at 15 lbs pressure (121°C) for 15 minutes. Mix well and pour into sterile Petri plates. If the solidified culture medium is somewhat too red, then to remove the colour, add a few drops (max. 1ml/litre) of a freshly prepared 10% Sodium sulphite solution and boil.

Caution: Basic fuchsin is a potential carcinogen and care should be taken to avoid inhalation of the powdered dye and contamination of the skin.

Quality Control

Physical Appearance

Light pink to purple homogeneous free flowing powder

Gelling

Firm, comparable with 1.2% Agar gel

Colour and Clarity of prepared medium

Orangish pink coloured, clear to slightly opalescent gel with fine precipitate forms in Petri plates.

Reaction

Reaction of 3.77% w/v aqueous solution at 25°C. pH : 7.5±0.2

pH range 7.30-7.70

Cultural Response/ characteristics

DM 2258: Cultural characteristics observed after an incubation at 35-37°C for 18-24 hours. pH 7.40-7.80

Organism	Inoculum (CFU)	Growth	Recovery	Colour of colony
<i>Bacillus subtilis</i> ATCC 6633	$\geq 10^3$	inhibited	0%	
<i>Enterobacter aerogenes</i> ATCC 13048	50-100	good-luxuriant	$\geq 50\%$	pink
<i>Enterococcus faecalis</i> ATCC 29212	50-100	none-poor	$\leq 10\%$	pink, small
<i>Escherichia coli</i> ATCC 25922	50-100	good-luxuriant	$\geq 50\%$	pink to rose red with metallic sheen
<i>Klebsiella pneumoniae</i> ATCC 13883	50-100	good-luxuriant	$\geq 50\%$	pink, mucoid
<i>Proteus vulgaris</i> ATCC 13315	50-100	good-luxuriant	$\geq 50\%$	colourless to pale pink
<i>Pseudomonas aeruginosa</i> ATCC 27853	50-100	good-luxuriant	$\geq 50\%$	colourless, irregular
<i>Salmonella Typhi</i> ATCC 6539	50-100	good-luxuriant	$\geq 50\%$	colourless to pale pink
<i>Shigella sonnei</i> ATCC 25931	50-100	good-luxuriant	$\geq 50\%$	irregular colourless to pale pink
<i>Staphylococcus aureus</i> ATCC 25923	$\geq 10^3$	inhibited	0%	pale pink
<i>Enterobacter cloacae</i> ATCC 13047	50-100	good	40-50%	pink
<i>Salmonella Typhimurium</i> ATCC 14028	50-100	good-luxuriant	$\geq 50\%$	colourless
<i>Salmonella Enteritidis</i> ATCC 13076	50-100	good-luxuriant	$\geq 50\%$	colourless
<i>Shigella flexneri</i> ATCC 12022	50-100	good-luxuriant	$\geq 50\%$	colourless

Storage and Shelf Life

Dried Media: Store below 30°C in tightly closed container and use before expiry date as mentioned on the label.

Prepared Media: 2-8°C in sealable plastic bags for 2-5 days.

Further Reading

1. Endo S., 1904, Centralbl. Bakt. I. Orig., 35:109.
2. Deutsches Fleischbeschaugesetz: Anlage Zu § 20 Abs. 4: Vorschriften über die bakteriologische Fleischuntersuchung.

Disclaimer :

- User must ensure suitability of the product(s) in their application prior to use.
- The product conform solely to the technical information provided in this booklet and to the best of knowledge research and development work carried at CDH is true and accurate
- **Central Drug House Pvt. Ltd.** reserves the right to make changes to specifications and information related to the products at any time.
- Products are not intended for human or animal diagnostic or therapeutic use but for laboratory, research or further manufacturing of diagnostic reagents extra.
- Statements contained herein should not be considered as a warranty of any kind, expressed or implied, and no liability is accepted for infringement of any patents.
- Do not use the products if it fails to meet specifications for identity and performance parameters.