

## Central Drug House (P) Ltd.

AN ISO 9001:2008 CERTIFIED COMPANY

Manufacturers of : Laboratory Fine Chemicals & Dehydrated Culture Media www.cdhfinechemical.com

<b>Product Specification</b>		1 A CI	• 📭	4 •
	Proc		neciti	cation
	TIUU	luct D		Caulon

## **INDICATOR PAPER** UNIVERSAL PH 1-10

(INCLUDING COLOUR SCALE) ACCURATE TO 1 pH UNIT

PRODUCT CODE	PA0786
SYNONYMS	<del>-</del>
C.I. NO.	<del></del>
CASR NO.	-
ATOMIC OR MOLECULAR FORMULA	<del></del>
ATOMIC OR MOLECULAR WEIGHT	-
PROPERTIES	

--

## PARAMETER LIMIT

Description Bright orangish yellow strips.

Sensitivity To pass the test.

## **pH-INDICATOR PAPERS:**

Possible discolouration of the dry new papers may be caused by their high sensitivity against and pH action. This does not impair the efficiency of the Indicator Papers for pH determination. With very weakly buffered or unbuffered solutions, the so-called indicator error occurs. This may be compensated for up to a point in the following manner: The strip is made to adhere to the inner wall of a test tube, which is then filled to the upper edge of the paper with the fluid to be tested. After 1/2 to 1 minute, the colour of the paper is compared with the scale through the glass of the test tube.

WARNING	IMDG Code :
Hazard statements: - Not hazardous. No hazards.	UN No. :
Precautionary statements	IATA :
Prevention:	
Responce:	
Disposal:	

Hazard Pictogram(s) :-----